

Sheet1

Lcal=Acal

Port J-number	Lcal (cold)	Lcal(hot)	ratio	Rcal(cold)	Rcal(hot)
1	-162.1	-138.2	0.85	145.9	127.1
2	-240.4	-214.1	0.89	156.7	138.2
3	-140.7	-121.7	0.86	128.2	112.6
4	-39.8	-26.9	0.68	43.6	32.4
5	121.7	114.3	0.94	-141.6	-124.6
6	228.9	217.9	0.95	-161.6	-146.8
7	145	153.5	1.06	-153.3	
8	16.6	15	0.9	-21	-17.4
9	54	21.6	0.4	-68.8	-41.9
10	246.5			-192.4	
11	250.7			-289.7	
12	217.2	162.2	0.75	-118.6	-86.9
13	-55.4	-28.7	0.52	67	44.4
14	-226.1	-195.4	0.86	168.4	143.8
15	-224.9	-180.7	0.8	267.5	222
16	-191.7	-152.7	0.8	130.7	97.9

Linearity results from Hot/Cold scans
columns are raw cal counts

worst 4	rf power
j9	-20.7
j13	-20.7
j4	-22.5
j12	-21.8

ratio	Delta(ratio)
0.87	0.02
0.88	-0.01
0.88	0.01
0.74	0.07
0.88	-0.06
0.91	-0.04
0.83	-0.08
0.61	0.21
0.73	-0.01
0.66	0.14
0.85	-0.01
0.83	0.03
0.75	-0.05